

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061223\
 Data File : VY014120.D
 Acq On : 12 Jun 2023 15:09
 Operator : KP/MD
 Sample : 03109-03
 Misc : 4.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-1-GRAB

Quant Time: Jun 13 05:42:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	156118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	265485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	269293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	108724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	113249	52.093	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.180%
35) Dibromofluoromethane	7.710	113	91686	48.698	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.400%
50) Toluene-d8	10.179	98	326275	46.353	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.700%
62) 4-Bromofluorobenzene	12.477	95	108401	43.353	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.700%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	19842	6.355	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061223\
Data File : VY014120.D
Acq On : 12 Jun 2023 15:09
Operator : KP/MD
Sample : 03109-03
Misc : 4.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-1-GRAB

Quant Time: Jun 13 05:42:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

