

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061223\
 Data File : VY014126.D
 Acq On : 12 Jun 2023 17:29
 Operator : KP/MD
 Sample : 03141-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CMC-EME-Topsoil-002

Quant Time: Jun 13 05:44:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	1807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	3083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	3025	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1763	70.064	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.120%	
35) Dibromofluoromethane	7.716	113	1037	47.430	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 94.860%	
50) Toluene-d8	10.179	98	3898	47.688	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 95.380%	
62) 4-Bromofluorobenzene	12.477	95	742	25.554	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 51.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061223\
Data File : VY014126.D
Acq On : 12 Jun 2023 17:29
Operator : KP/MD
Sample : 03141-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CMC-EME-Topsoil-002

Quant Time: Jun 13 05:44:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

