

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061323\
 Data File : VY014145.D
 Acq On : 13 Jun 2023 17:26
 Operator : KP/MD
 Sample : 03167-01
 Misc : 5.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-4-GA

Quant Time: Jun 14 01:37:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	159691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	280378	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	298279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	153525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	114360	51.427	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 102.860%			
35) Dibromofluoromethane	7.716	113	96474	48.519	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 97.040%			
50) Toluene-d8	10.179	98	351889	47.337	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 94.680%			
62) 4-Bromofluorobenzene	12.477	95	134180	50.813	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 101.620%			
Target Compounds						
16) Acetone	3.948	43	13673	20.398	ug/l	98
25) 2-Butanone	6.990	43	7347	8.904	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061323\
Data File : VY014145.D
Acq On : 13 Jun 2023 17:26
Operator : KP/MD
Sample : 03167-01
Misc : 5.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-4-GA

Quant Time: Jun 14 01:37:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

