

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY061820\
 Data File : VY002919.D
 Acq On : 18 Jun 2020 15:12
 Operator : SY/MD
 Sample : L3041-01
 Misc : 4.99G/5ML/MSVOA Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PAD

Quant Time: Jun 19 05:32:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 13:13:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	228361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	372447	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	355457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	177936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	112826	56.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.80%	
35) Dibromofluoromethane	7.72	113	111752	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	10.18	98	441349	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	12.48	95	148951	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
Target Compounds						
3) Chloromethane	2.11	50	2765	1.591	ug/l	89
16) Acetone	3.96	43	4755	10.227	ug/l	91
20) Methylene Chloride	4.72	84	30420	13.388	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY061820\
Data File : VY002919.D
Acq On : 18 Jun 2020 15:12
Operator : SY/MD
Sample : L3041-01
Misc : 4.99G/5ML/MSVOA Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PAD

Quant Time: Jun 19 05:32:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
Quant Title : SW846 8260
QLast Update : Wed May 27 13:13:33 2020
Response via : Initial Calibration

