

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY061820\
 Data File : VY002921.D
 Acq On : 18 Jun 2020 16:00
 Operator : SY/MD
 Sample : L3037-01
 Misc : 7.38G/5ML/MSVOA Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SL-WC-01

Quant Time: Jun 19 05:40:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 13:13:33 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	240009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	393480	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	363280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	166110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	121976	58.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.06%	
35) Dibromofluoromethane	7.72	113	121891	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
50) Toluene-d8	10.18	98	461580	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	12.48	95	148504	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
Target Compounds						
3) Chloromethane	2.12	50	2166	1.186	ug/l	99
20) Methylene Chloride	4.73	84	20587	8.621	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY061820\
Data File : VY002921.D
Acq On : 18 Jun 2020 16:00
Operator : SY/MD
Sample : L3037-01
Misc : 7.38G/5ML/MSVOA Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SL-WC-01

Quant Time: Jun 19 05:40:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y052720S.M
Quant Title : SW846 8260
QLast Update : Wed May 27 13:13:33 2020
Response via : Initial Calibration

