


Tert butyl alcohol


Response Ratio



Response = 4.980e-002 * Amt + 8.232e-002
 Coef of Det (r^2) = 0.996356 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y062422S.M
 Calibration Table Last Updated: Sat Jun 25 07:08:46 2022

Toluene-d8

Response Ratio



$$\text{Response} = 6.814\text{e-}001 * \text{Amt} + 1.071\text{e-}001$$


Coef of Det (r^2) = 0.995361 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y062422S.M

Calibration Table Last Updated: Sat Jun 25 07:08:46 2022

1,2-Dichloroethane-d4

Response Ratio



$$\text{Response} = 3.419\text{e-}001 * \text{Amt} + 3.468\text{e-}002$$


Coef of Det (r^2) = 0.997274 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y062422S.M

Calibration Table Last Updated: Sat Jun 25 07:08:46 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 5.122\text{e-}001 * \text{Amt} + 3.760\text{e-}002$$

Coef of Det (r^2) = 0.998757 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y062422S.M

Calibration Table Last Updated: Sat Jun 25 07:08:46 2022