

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009321.D
 Acq On : 25 Jun 2022 14:11
 Operator : KP/MD
 Sample : N3488-01
 Misc : 3.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-9-0-2-20220623

Quant Time: Jun 27 10:27:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	179827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	306172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	286606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	128504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	105010	80.324	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 160.640%	
35) Dibromofluoromethane	7.716	113	94310	57.068	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.140%	
50) Toluene-d8	10.173	98	366393	79.949	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 159.900%#	
62) 4-Bromofluorobenzene	12.477	95	121219	51.503	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009321.D
Acq On : 25 Jun 2022 14:11
Operator : KP/MD
Sample : N3488-01
Misc : 3.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-9-0-2-20220623

Quant Time: Jun 27 10:27:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

