

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009324.D
 Acq On : 25 Jun 2022 15:21
 Operator : KP/MD
 Sample : N3488-04
 Misc : 3.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10-0-2-20220623

Quant Time: Jun 27 10:29:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	171523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	300248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.484	117	280427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	104765	84.250	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 168.500%#		
35) Dibromofluoromethane	7.710	113	91395	56.395	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 112.800%		
50) Toluene-d8	10.173	98	360934	80.347	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 160.700%#		
62) 4-Bromofluorobenzene	12.477	95	118310	51.259	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 102.520%		
Target Compounds						
16) Acetone	3.948	43	4982	11.361	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009324.D
Acq On : 25 Jun 2022 15:21
Operator : KP/MD
Sample : N3488-04
Misc : 3.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10-0-2-20220623

Quant Time: Jun 27 10:29:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

