

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009325.D
 Acq On : 25 Jun 2022 15:45
 Operator : KP/MD
 Sample : N3488-05
 Misc : 4.57g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10-6-8-20220623

Quant Time: Jun 27 10:29:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	169572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	296489	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	279024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	130153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	103019	83.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 167.540%#	
35) Dibromofluoromethane	7.716	113	92160	57.588	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.180%	
50) Toluene-d8	10.173	98	354295	79.823	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 159.640%#	
62) 4-Bromofluorobenzene	12.477	95	120598	52.912	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.820%	
Target Compounds						
						Qvalue
16) Acetone	3.936	43	46848	108.064	ug/l	98
17) Carbon Disulfide	4.180	76	15356	2.982	ug/l	96
25) 2-Butanone	6.984	43	17209	26.397	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009325.D
Acq On : 25 Jun 2022 15:45
Operator : KP/MD
Sample : N3488-05
Misc : 4.57g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10-6-8-20220623

Quant Time: Jun 27 10:29:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

