

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009330.D
 Acq On : 25 Jun 2022 17:42
 Operator : KP/MD
 Sample : N3488-10
 Misc : 5.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-12-17-19-20220623

Quant Time: Jun 27 10:31:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	161201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	277812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	267111	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	119931	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	94170	80.357	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 160.720%	
35) Dibromofluoromethane	7.710	113	84260	56.192	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.380%	
50) Toluene-d8	10.173	98	334018	80.362	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 160.720%#	
62) 4-Bromofluorobenzene	12.477	95	111365	52.146	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009330.D
Acq On : 25 Jun 2022 17:42
Operator : KP/MD
Sample : N3488-10
Misc : 5.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-12-17-19-20220623

Quant Time: Jun 27 10:31:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

