

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009331.D
 Acq On : 25 Jun 2022 18:05
 Operator : KP/MD
 Sample : N3488-11
 Misc : 4.55g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-7-0-2-20220623

Quant Time: Jun 27 10:32:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	158072	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	274090	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	256109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	116856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	95693	83.457	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 166.920%#	
35) Dibromofluoromethane	7.710	113	84771	57.300	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.600%	
50) Toluene-d8	10.173	98	330046	80.496	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 161.000%#	
62) 4-Bromofluorobenzene	12.477	95	107745	51.136	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.280%	
Target Compounds						
16) Acetone	3.942	43	5371	13.291	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009331.D
Acq On : 25 Jun 2022 18:05
Operator : KP/MD
Sample : N3488-11
Misc : 4.55g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-7-0-2-20220623

Quant Time: Jun 27 10:32:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

