

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009334.D
 Acq On : 25 Jun 2022 19:15
 Operator : KP/MD
 Sample : N3488-14
 Misc : 4.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-13-15-17-20220623

Quant Time: Jun 27 10:33:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	151719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	267307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	232264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	115061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	91139	82.775	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 165.540%#	
35) Dibromofluoromethane	7.710	113	79831	55.330	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.660%	
50) Toluene-d8	10.173	98	320912	80.231	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 160.460%#	
62) 4-Bromofluorobenzene	12.477	95	107094	52.117	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.240%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009334.D
Acq On : 25 Jun 2022 19:15
Operator : KP/MD
Sample : N3488-14
Misc : 4.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-13-15-17-20220623

Quant Time: Jun 27 10:33:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

