

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009335.D
 Acq On : 25 Jun 2022 19:39
 Operator : KP/MD
 Sample : N3488-15
 Misc : 3.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-14-0-2-20220623

Quant Time: Jun 27 10:33:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	29230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	44990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	40720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	16748	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	17595	82.956	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	165.920%#
35) Dibromofluoromethane	7.716	113	15226	62.701	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.400%
50) Toluene-d8	10.173	98	54185	80.513	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	161.020%#
62) 4-Bromofluorobenzene	12.477	95	15539	44.930	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009335.D
Acq On : 25 Jun 2022 19:39
Operator : KP/MD
Sample : N3488-15
Misc : 3.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-14-0-2-20220623

Quant Time: Jun 27 10:33:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

