

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
 Data File : VY009336.D
 Acq On : 25 Jun 2022 20:02
 Operator : KP/MD
 Sample : N3488-16
 Misc : 4.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-14-2.5-4.5-20220623

Quant Time: Jun 27 10:34:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 25 07:08:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	153472	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	270841	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	259899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	103833	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	93303	83.834	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	167.660%#
35) Dibromofluoromethane	7.710	113	82198	56.227	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.460%
50) Toluene-d8	10.173	98	325028	80.197	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	160.400%#
62) 4-Bromofluorobenzene	12.477	95	108325	52.028	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062522\
Data File : VY009336.D
Acq On : 25 Jun 2022 20:02
Operator : KP/MD
Sample : N3488-16
Misc : 4.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-14-2.5-4.5-20220623

Quant Time: Jun 27 10:34:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062422S.M
Quant Title : SW846 8260
QLast Update : Sat Jun 25 07:08:46 2022
Response via : Initial Calibration

