

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062923\
 Data File : VY014496.D
 Acq On : 29 Jun 2023 15:15
 Operator : KP/MD
 Sample : 03473-02
 Misc : 5.58g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CMC-EME-TOPSOIL-006

Quant Time: Jun 30 04:09:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	178317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	248306	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	230888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	106199	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	128472	58.028	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.060%
35) Dibromofluoromethane	7.709	113	77488	50.511	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.020%
50) Toluene-d8	10.179	98	288213	47.262	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.520%
62) 4-Bromofluorobenzene	12.477	95	89552	41.134	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062923\
Data File : VY014496.D
Acq On : 29 Jun 2023 15:15
Operator : KP/MD
Sample : 03473-02
Misc : 5.58g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CMC-EME-TOPSOIL-006

Quant Time: Jun 30 04:09:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

