

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062923\
 Data File : VY014506.D
 Acq On : 29 Jun 2023 19:11
 Operator : KP/MD
 Sample : 03473-12
 Misc : 5.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CMC-EME-TOPSOIL-017

Quant Time: Jun 30 04:12:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 17:18:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	191371	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	267829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	252365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	122868	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	136186	57.316	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.640%
35) Dibromofluoromethane	7.716	113	80553	48.681	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.360%
50) Toluene-d8	10.179	98	308607	46.918	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.840%
62) 4-Bromofluorobenzene	12.477	95	100064	42.612	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062923\
Data File : VY014506.D
Acq On : 29 Jun 2023 19:11
Operator : KP/MD
Sample : 03473-12
Misc : 5.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CMC-EME-TOPSOIL-017

Quant Time: Jun 30 04:12:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062323S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 17:18:46 2023
Response via : Initial Calibration

