

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063023\
 Data File : VY014524.D
 Acq On : 30 Jun 2023 18:40
 Operator : KP/MD
 Sample : 03486-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BAYONNE-COMP

Quant Time: Jun 30 23:20:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 30 19:02:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	288302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	454205	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	439036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	211945	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	163189	56.900	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.800%
35) Dibromofluoromethane	7.710	113	148121	52.957	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.920%
50) Toluene-d8	10.179	98	537727	48.632	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.260%
62) 4-Bromofluorobenzene	12.477	95	177868	46.327	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063023\
Data File : VY014524.D
Acq On : 30 Jun 2023 18:40
Operator : KP/MD
Sample : 03486-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BAYONNE-COMP

Quant Time: Jun 30 23:20:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y063023S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 30 19:02:11 2023
Response via : Initial Calibration

