

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071822\
 Data File : VY009577.D
 Acq On : 18 Jul 2022 16:50
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jul 19 04:25:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	146851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	247900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	231296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	110956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	57029	41.062	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.120%
35) Dibromofluoromethane	7.710	113	52896	33.869	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	67.740%
50) Toluene-d8	10.172	98	198353	33.038	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	66.080%#
62) 4-Bromofluorobenzene	12.477	95	64431	33.119	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	66.240%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	35247	12.617	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071822\
Data File : VY009577.D
Acq On : 18 Jul 2022 16:50
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jul 19 04:25:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

