

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071922\
 Data File : VY009590.D
 Acq On : 19 Jul 2022 15:28
 Operator : KP/MD
 Sample : N3651-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SED-1-(6-12)

Quant Time: Jul 20 02:44:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	195418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	332568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	307033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	123440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	92010	49.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.560%
35) Dibromofluoromethane	7.710	113	87012	41.530	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.060%
50) Toluene-d8	10.173	98	330727	41.061	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	82.120%
62) 4-Bromofluorobenzene	12.477	95	100438	38.484	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	76.960%
Target Compounds						
17) Carbon Disulfide	4.180	76	73268	13.115	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071922\
Data File : VY009590.D
Acq On : 19 Jul 2022 15:28
Operator : KP/MD
Sample : N3651-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SED-1-(6-12)

Quant Time: Jul 20 02:44:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

