

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071922\
 Data File : VY009599.D
 Acq On : 19 Jul 2022 19:00
 Operator : KP/MD
 Sample : N3652-02
 Misc : 4.76g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SED-7-(6-12)

Quant Time: Jul 20 02:45:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	277638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	467318	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	447016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	214129	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	123177	46.911	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.820%
35) Dibromofluoromethane	7.710	113	120222	40.835	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.680%
50) Toluene-d8	10.173	98	468349	41.381	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	82.760%
62) 4-Bromofluorobenzene	12.477	95	167386	45.642	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	91.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071922\
Data File : VY009599.D
Acq On : 19 Jul 2022 19:00
Operator : KP/MD
Sample : N3652-02
Misc : 4.76g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SED-7-(6-12)

Quant Time: Jul 20 02:45:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

