

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv072120\
 Data File : VY003253.D
 Acq On : 21 Jul 2020 16:27
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 22 03:36:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 13:18:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.358	0.287	19.8	87	0.00
3 P	Chloromethane	0.497	0.443	10.9	93	0.00
4 C	Vinyl Chloride	0.510	0.472	7.5#	94	0.00
5 T	Bromomethane	0.379	0.362	4.5	101	0.00
6 T	Chloroethane	0.335	0.313	6.6	94	0.00
7 T	Trichlorofluoromethane	0.728	0.698	4.1	97	0.00
8 T	Diethyl Ether	0.244	0.250	-2.5	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.440	0.428	2.7	97	0.00
10 T	Methyl Iodide	0.478	0.456	4.6	89	0.00
11 T	Tert butyl alcohol	0.046	0.046	0.0	107	0.00
12 CM	1,1-Dichloroethene	0.416	0.409	1.7#	98	0.00
13 T	Acrolein	0.040	0.037	7.5	98	0.00
14 T	Allyl chloride	0.617	0.668	-8.3	106	0.00
15 T	Acrylonitrile	0.117	0.131	-12.0	110	0.00
16 T	Acetone	0.094	0.108	-14.9	111	0.00
17 T	Carbon Disulfide	1.277	1.229	3.8	95	0.00
18 T	Methyl Acetate	0.262	0.297	-13.4	116	0.00
19 T	Methyl tert-butyl Ether	1.166	1.228	-5.3	103	0.00
20 T	Methylene Chloride	0.560	0.494	11.8	98	0.00
21 T	trans-1,2-Dichloroethene	0.471	0.467	0.8	98	0.00
22 T	Diisopropyl ether	1.370	1.506	-9.9	106	0.00
23 T	Vinyl Acetate	0.913	1.005	-10.1	105	0.00
24 P	1,1-Dichloroethane	0.794	0.833	-4.9	103	0.00
25 T	2-Butanone	0.155	0.172	-11.0	109	0.00
26 T	2,2-Dichloropropane	0.715	0.717	-0.3	99	0.00
27 T	cis-1,2-Dichloroethene	0.518	0.533	-2.9	101	0.00
28 T	Bromochloromethane	0.344	0.360	-4.7	101	0.00
29 T	Tetrahydrofuran	0.104	0.116	-11.5	109	0.00
30 C	Chloroform	0.818	0.852	-4.2#	103	0.00
31 T	Cyclohexane	0.768	0.734	4.4	96	0.00
32 T	1,1,1-Trichloroethane	0.736	0.739	-0.4	99	0.00
33 S	1,2-Dichloroethane-d4	0.422	0.424	-0.5	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.288	0.278	3.5	96	0.00
36 T	1,1-Dichloropropene	0.453	0.439	3.1	97	0.00
37 T	Ethyl Acetate	0.244	0.266	-9.0	108	0.00
38 T	Carbon Tetrachloride	0.468	0.457	2.4	97	0.00
39 T	Methylcyclohexane	0.569	0.547	3.9	97	0.00
40 TM	Benzene	1.248	1.262	-1.1	102	0.00
41 T	Methacrylonitrile	0.147	0.139	5.4	92	0.00
42 TM	1,2-Dichloroethane	0.384	0.388	-1.0	102	0.00
43 T	Isopropyl Acetate	0.472	0.503	-6.6	106	0.00
44 TM	Trichloroethene	0.383	0.369	3.7	98	0.00
45 C	1,2-Dichloropropane	0.323	0.335	-3.7#	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv072120\
 Data File : VY003253.D
 Acq On : 21 Jul 2020 16:27
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 22 03:36:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 13:18:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.189	0.193	-2.1	102	0.00
47 T	Bromodichloromethane	0.448	0.463	-3.3	103	0.00
48 T	Methyl methacrylate	0.211	0.225	-6.6	105	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	101	0.00
50 S	Toluene-d8	1.070	1.019	4.8	96	0.00
51 T	4-Methyl-2-Pentanone	0.246	0.266	-8.1	107	0.00
52 CM	Toluene	0.798	0.791	0.9#	100	0.00
53 T	t-1,3-Dichloropropene	0.465	0.484	-4.1	102	0.00
54 T	cis-1,3-Dichloropropene	0.535	0.555	-3.7	104	0.00
55 T	1,1,2-Trichloroethane	0.267	0.271	-1.5	103	0.00
56 T	Ethyl methacrylate	0.363	0.390	-7.4	104	0.00
57 T	1,3-Dichloropropane	0.453	0.474	-4.6	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.185	0.197	-6.5	105	0.00
59 T	2-Hexanone	0.171	0.188	-9.9	108	0.00
60 T	Dibromochloromethane	0.336	0.348	-3.6	103	0.00
61 T	1,2-Dibromoethane	0.264	0.263	0.4	101	0.00
62 S	4-Bromofluorobenzene	0.386	0.369	4.4	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.446	0.407	8.7	95	0.00
65 PM	Chlorobenzene	0.963	0.950	1.3	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.372	0.5	102	0.00
67 C	Ethyl Benzene	1.720	1.694	1.5#	100	0.00
68 T	m/p-Xylenes	0.653	0.640	2.0	101	0.00
69 T	o-Xylene	0.610	0.614	-0.7	102	0.00
70 T	Styrene	1.043	1.068	-2.4	103	0.00
71 P	Bromoform	0.246	0.255	-3.7	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.186	3.129	1.8	100	0.00
74 T	N-amyl acetate	0.894	0.956	-6.9	106	0.00
75 P	1,1,2,2-Tetrachloroethane	0.615	0.643	-4.6	107	0.00
76 T	1,2,3-Trichloropropane	0.443	0.470	-6.1	109	0.00
77 T	Bromobenzene	0.812	0.809	0.4	102	0.00
78 T	n-propylbenzene	3.756	3.750	0.2	101	0.00
79 T	2-Chlorotoluene	2.096	2.082	0.7	101	0.00
80 T	1,3,5-Trimethylbenzene	2.656	2.628	1.1	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.227	0.237	-4.4	105	0.00
82 T	4-Chlorotoluene	2.187	2.189	-0.1	102	0.00
83 T	tert-Butylbenzene	2.343	2.345	-0.1	101	0.00
84 T	1,2,4-Trimethylbenzene	2.672	2.657	0.6	101	0.00
85 T	sec-Butylbenzene	3.221	3.219	0.1	101	0.00
86 T	p-Isopropyltoluene	3.007	2.985	0.7	100	0.00
87 T	1,3-Dichlorobenzene	1.536	1.545	-0.6	103	0.00
88 T	1,4-Dichlorobenzene	1.521	1.531	-0.7	104	0.00
89 T	n-Butylbenzene	2.829	2.812	0.6	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv072120\
Data File : VY003253.D
Acq On : 21 Jul 2020 16:27
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 22 03:36:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 13:18:03 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.548	0.561	-2.4	102	0.00
91 T	1,2-Dichlorobenzene	1.348	1.396	-3.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.110	-10.0	115	0.00
93 T	1,2,4-Trichlorobenzene	0.959	1.004	-4.7	107	0.00
94 T	Hexachlorobutadiene	0.657	0.638	2.9	101	0.00
95 T	Naphthalene	1.585	1.898	-19.7	123	0.00
96 T	1,2,3-Trichlorobenzene	0.758	0.873	-15.2	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6