

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072622\
 Data File : VY009699.D
 Acq On : 26 Jul 2022 14:04
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jul 27 04:25:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 27 04:24:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	173447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	280017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	260914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	121952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	89816	80.998	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 162.000%	
35) Dibromofluoromethane	7.710	113	86488	56.790	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.580%	
50) Toluene-d8	10.179	98	322192	74.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 149.480%#	
62) 4-Bromofluorobenzene	12.477	95	105494	52.241	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 104.480%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.704	84	34024	9.186	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072622\
Data File : VY009699.D
Acq On : 26 Jul 2022 14:04
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jul 27 04:25:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072622S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 27 04:24:58 2022
Response via : Initial Calibration

