

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
 Data File : VY009965.D
 Acq On : 11 Aug 2022 15:45
 Operator : KP/MD
 Sample : N4119-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP1-2-4FT

Quant Time: Aug 12 02:49:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 17:37:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.416	152	87	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	107	34.555	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	69.120%
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.185	98	564	44.143	ug/l	0.01
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.280%
62) 4-Bromofluorobenzene	12.483	95	19	3.140	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	6.280%#
Target Compounds						Qvalue
10) Methyl Iodide	4.088	142	21	4.134	ug/l	# 40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
Data File : VY009965.D
Acq On : 11 Aug 2022 15:45
Operator : KP/MD
Sample : N4119-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP1-2-4FT

Quant Time: Aug 12 02:49:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 17:37:28 2022
Response via : Initial Calibration

