

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
 Data File : VY009975.D
 Acq On : 11 Aug 2022 19:41
 Operator : KP/MD
 Sample : N4064-07
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-15

Quant Time: Aug 12 02:50:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 17:37:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	65809	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	135036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	141823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	29058	57.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.280%
35) Dibromofluoromethane	7.710	113	27254	35.078	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	70.160%
50) Toluene-d8	10.173	98	98625	42.530	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.060%
62) 4-Bromofluorobenzene	12.477	95	32354	29.620	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	59.240%
Target Compounds						
16) Acetone	3.942	43	3133	18.924	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
Data File : VY009975.D
Acq On : 11 Aug 2022 19:41
Operator : KP/MD
Sample : N4064-07
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-15

Quant Time: Aug 12 02:50:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 17:37:28 2022
Response via : Initial Calibration

