

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
 Data File : VY009978.D
 Acq On : 11 Aug 2022 20:52
 Operator : KP/MD
 Sample : N4119-01RE
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP1-2-4FT

Quant Time: Aug 12 02:51:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 17:37:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	43073	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	85768	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	83537	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	32873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	16854	50.351	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.700%
35) Dibromofluoromethane	7.710	113	16683	33.807	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	67.620%
50) Toluene-d8	10.173	98	61839	41.892	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	83.780%
62) 4-Bromofluorobenzene	12.477	95	17229	24.834	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	49.660%
Target Compounds						
16) Acetone	3.948	43	4375	40.374	ug/l #	82
20) Methylene Chloride	4.698	84	5981	7.502	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
Data File : VY009978.D
Acq On : 11 Aug 2022 20:52
Operator : KP/MD
Sample : N4119-01RE
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP1-2-4FT

Quant Time: Aug 12 02:51:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 17:37:28 2022
Response via : Initial Calibration

