

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
 Data File : VY009979.D
 Acq On : 11 Aug 2022 21:16
 Operator : KP/MD
 Sample : N4127-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-080922

Quant Time: Aug 12 02:51:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 17:37:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	64284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	135005	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	141488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	64000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	28728	58.411	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.820%
35) Dibromofluoromethane	7.710	113	26518	34.139	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	68.280%
50) Toluene-d8	10.173	98	99908	43.190	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.380%
62) 4-Bromofluorobenzene	12.477	95	31792	29.113	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	58.220%
Target Compounds						
16) Acetone	3.942	43	5841	36.117	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081122\
Data File : VY009979.D
Acq On : 11 Aug 2022 21:16
Operator : KP/MD
Sample : N4127-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-080922

Quant Time: Aug 12 02:51:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 17:37:28 2022
Response via : Initial Calibration

