

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
 Data File : VY010501.D
 Acq On : 15 Sep 2022 11:24
 Operator : KP/MD
 Sample : VY0915SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0915SBL01

Quant Time: Sep 15 12:08:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	217402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	351053	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	310112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	150099	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	117572	53.696	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.400%
35) Dibromofluoromethane	7.716	113	116238	46.622	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.240%
50) Toluene-d8	10.179	98	411813	50.965	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.940%
62) 4-Bromofluorobenzene	12.483	95	143290	44.000	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
Data File : VY010501.D
Acq On : 15 Sep 2022 11:24
Operator : KP/MD
Sample : VY0915SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0915SBL01

Quant Time: Sep 15 12:08:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

