

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
 Data File : VY010505.D
 Acq On : 15 Sep 2022 12:57
 Operator : KP/MD
 Sample : N4588-02
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 40152

Quant Time: Sep 15 16:03:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	191974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	331318	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	318210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	147081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	112112	57.984	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.960%
35) Dibromofluoromethane	7.716	113	103716	44.077	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.160%
50) Toluene-d8	10.179	98	398063	52.198	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.483	95	137192	44.738	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.480%
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	3266	7.361	ug/l	95
20) Methylene Chloride	4.710	84	23940	4.953	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
Data File : VY010505.D
Acq On : 15 Sep 2022 12:57
Operator : KP/MD
Sample : N4588-02
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
40152

Quant Time: Sep 15 16:03:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

