

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
 Data File : VY010511.D
 Acq On : 15 Sep 2022 15:16
 Operator : KP/MD
 Sample : N4630-04
 Misc : 3.38g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-17-4-1-GR

Quant Time: Sep 16 02:56:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	205554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	353730	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	296284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	118599	57.287	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.580%
35) Dibromofluoromethane	7.710	113	108900	43.348	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.700%
50) Toluene-d8	10.179	98	409951	50.351	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.700%
62) 4-Bromofluorobenzene	12.477	95	115652	33.847	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	67.700%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	7834	1.226	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
Data File : VY010511.D
Acq On : 15 Sep 2022 15:16
Operator : KP/MD
Sample : N4630-04
Misc : 3.38g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-17-4-1-GR

Quant Time: Sep 16 02:56:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

