

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
 Data File : VY010512.D
 Acq On : 15 Sep 2022 15:39
 Operator : KP/MD
 Sample : N4634-04
 Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2

Quant Time: Sep 16 02:56:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	204100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	349402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	331925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	157364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121165	58.943	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.880%
35) Dibromofluoromethane	7.710	113	108735	43.819	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.640%
50) Toluene-d8	10.179	98	416423	51.779	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.560%
62) 4-Bromofluorobenzene	12.477	95	147715	45.824	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.640%
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	11671	24.740	ug/l	90
18) Methyl Acetate	4.479	43	2881	1.985	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
Data File : VY010512.D
Acq On : 15 Sep 2022 15:39
Operator : KP/MD
Sample : N4634-04
Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2

Quant Time: Sep 16 02:56:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

