

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
 Data File : VY010513.D
 Acq On : 15 Sep 2022 16:02
 Operator : KP/MD
 Sample : N4634-01
 Misc : 4.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Sep 16 02:56:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	205843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	352073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	328220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	144824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	121152	58.438	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.880%
35) Dibromofluoromethane	7.710	113	109167	43.659	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.320%
50) Toluene-d8	10.179	98	417557	51.526	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.477	95	142172	43.455	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.900%
Target Compounds						
16) Acetone	3.948	43	6192	13.015	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
Data File : VY010513.D
Acq On : 15 Sep 2022 16:02
Operator : KP/MD
Sample : N4634-01
Misc : 4.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Sep 16 02:56:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

