

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
 Data File : VY010519.D
 Acq On : 15 Sep 2022 18:23
 Operator : KP/MD
 Sample : N4630-24
 Misc : 4.83g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-17-4-5-GR

Quant Time: Sep 16 02:57:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	185645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	319350	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	276092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	86011	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	109277	58.445	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.880%
35) Dibromofluoromethane	7.716	113	101026	44.543	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.080%
50) Toluene-d8	10.179	98	373441	50.804	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.600%
62) 4-Bromofluorobenzene	12.477	95	102841	33.233	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091522\
Data File : VY010519.D
Acq On : 15 Sep 2022 18:23
Operator : KP/MD
Sample : N4630-24
Misc : 4.83g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-17-4-5-GR

Quant Time: Sep 16 02:57:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

