

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091622\
 Data File : VY010525.D
 Acq On : 16 Sep 2022 11:02
 Operator : KP/MD
 Sample : VY0916SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0916SBL01

Quant Time: Sep 17 00:27:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	200839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	324423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	287096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	142739	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	111020	54.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.760%
35) Dibromofluoromethane	7.716	113	108819	47.229	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.460%
50) Toluene-d8	10.179	98	380885	51.007	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.020%
62) 4-Bromofluorobenzene	12.483	95	134180	44.677	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091622\
Data File : VY010525.D
Acq On : 16 Sep 2022 11:02
Operator : KP/MD
Sample : VY0916SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0916SBL01

Quant Time: Sep 17 00:27:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

