

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091622\
 Data File : VY010531.D
 Acq On : 16 Sep 2022 13:18
 Operator : KP/MD
 Sample : N4509-07
 Misc : 5.39g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S1-SO_PW002-20220902

Quant Time: Sep 17 00:28:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	195401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	314376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	277461	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133565	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	106772	54.254	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.500%
35) Dibromofluoromethane	7.716	113	103546	46.377	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.760%
50) Toluene-d8	10.179	98	366016	50.582	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.483	95	124461	42.465	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.940%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091622\
Data File : VY010531.D
Acq On : 16 Sep 2022 13:18
Operator : KP/MD
Sample : N4509-07
Misc : 5.39g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_PW002-20220902

Quant Time: Sep 17 00:28:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY010531.D
Acq: 16 Sep 2022 13:18




Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_PW002-20220902


Tgt Ion:168 Resp: 195401
Ion Ratio Lower Upper
168 100
99 55.2 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 54.254 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY010531.D
Acq: 16 Sep 2022 13:18

Tgt Ion: 65 Resp: 106772
Ion Ratio Lower Upper
65 100
67 53.1 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VY010531.D
Acq: 16 Sep 2022 13:18




Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_PW002-20220902


Tgt Ion:114 Resp: 314376
Ion Ratio Lower Upper
114 100
63 18.3 0.0 43.2
88 15.3 0.0 31.8



#35
Dibromofluoromethane
Concen: 46.377 ug/l
RT: 7.716 min Scan# 967
Delta R.T. 0.006 min
Lab File: VY010531.D
Acq: 16 Sep 2022 13:18

Tgt Ion:113 Resp: 103546
Ion Ratio Lower Upper
113 100
111 104.3 82.6 124.0
192 19.6 17.6 26.4









#50
Toluene-d8
Concen: 50.582 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY010531.D
Acq: 16 Sep 2022 13:18




Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_PW002-20220902


Tgt Ion: 98 Resp: 366016
Ion Ratio Lower Upper
98 100
100 65.0 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 42.465 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.006 min
Lab File: VY010531.D
Acq: 16 Sep 2022 13:18

Tgt Ion: 95 Resp: 124461
Ion Ratio Lower Upper
95 100
174 85.4 0.0 177.0
176 82.7 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.001 min

Lab File: VY010531.D

Acq: 16 Sep 2022 13:18

Instrument :

MSVOA_Y

ClientSampleId :

S1-SO_PW002-20220902





Tgt Ion:117 Resp: 277461

Ion Ratio Lower Upper

117 100

82 54.3 43.0 64.4

119 32.1 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY010531.D

Acq: 16 Sep 2022 13:18

Tgt Ion:152 Resp: 133565

Ion Ratio Lower Upper

152 100

115 58.0 28.2 84.7

150 157.4 0.0 347.6

