

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092822\
 Data File : VY010722.D
 Acq On : 28 Sep 2022 17:22
 Operator : KP/MD
 Sample : N4812-01RE
 Misc : 6.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-3RE

Quant Time: Sep 29 01:14:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	1398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	2462	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	425	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	901	67.382	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.760%
35) Dibromofluoromethane	7.722	113	370	23.548	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	47.100%#
50) Toluene-d8	10.179	98	2185	38.400	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	76.800%
62) 4-Bromofluorobenzene	12.477	95	566	24.907	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	49.820%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.710	84	600	38.623	ug/l	# 76
73) Isopropylbenzene	12.325	105	1631	62.096	ug/l	96
78) n-propylbenzene	12.666	91	885	28.287	ug/l	80
80) 1,3,5-Trimethylbenzene	12.812	105	880	39.652	ug/l	92
84) 1,2,4-Trimethylbenzene	13.117	105	1232	55.878	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092822\
Data File : VY010722.D
Acq On : 28 Sep 2022 17:22
Operator : KP/MD
Sample : N4812-01RE
Misc : 6.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-3RE

Quant Time: Sep 29 01:14:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

