

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
 Data File : VY010811.D
 Acq On : 05 Oct 2022 11:52
 Operator : KP/MD
 Sample : N4948-01
 Misc : 6.34g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-6

Quant Time: Oct 06 09:16:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	198248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	338892	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	292189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	128182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	100032	49.780	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.560%
35) Dibromofluoromethane	7.716	113	104346	46.384	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.760%
50) Toluene-d8	10.179	98	330751	42.992	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.980%
62) 4-Bromofluorobenzene	12.477	95	119170	38.826	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.660%
Target Compounds						
16) Acetone	3.948	43	6222	12.191	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
Data File : VY010811.D
Acq On : 05 Oct 2022 11:52
Operator : KP/MD
Sample : N4948-01
Misc : 6.34g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-6

Quant Time: Oct 06 09:16:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

