

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
 Data File : VY010819.D
 Acq On : 05 Oct 2022 14:53
 Operator : KP/MD
 Sample : N4960-05
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PAT-03-100322

Quant Time: Oct 06 09:18:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	238652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	410967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	348025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	144395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	123450	51.033	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.060%	
35) Dibromofluoromethane	7.710	113	131986	48.381	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 96.760%	
50) Toluene-d8	10.179	98	406257	43.545	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 87.100%	
62) 4-Bromofluorobenzene	12.477	95	146522	39.421	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 78.840%	
Target Compounds						Qvalue
16) Acetone	3.942	43	1714	2.790	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
Data File : VY010819.D
Acq On : 05 Oct 2022 14:53
Operator : KP/MD
Sample : N4960-05
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PAT-03-100322

Quant Time: Oct 06 09:18:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

