

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
 Data File : VY010826.D
 Acq On : 05 Oct 2022 17:32
 Operator : KP/MD
 Sample : N5004-06
 Misc : 6.64g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DA-21-17

Quant Time: Oct 06 09:22:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	164576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	301013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	214443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	55312	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	117904	70.678	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 141.360%	
35) Dibromofluoromethane	7.716	113	121765	60.939	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 121.880%	
50) Toluene-d8	10.179	98	315867	46.224	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.440%	
62) 4-Bromofluorobenzene	12.477	95	77937	27.540	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 55.080%	
Target Compounds						Qvalue
16) Acetone	3.936	43	5575	13.159	ug/l	# 88
19) Methyl tert-butyl Ether	5.216	73	31950	7.844	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
Data File : VY010826.D
Acq On : 05 Oct 2022 17:32
Operator : KP/MD
Sample : N5004-06
Misc : 6.64g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DA-21-17

Quant Time: Oct 06 09:22:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

