



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100522\
Data File : VY010806.D
Acq On : 05 Oct 2022 09:15
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Title : SW846 8260
Last Update : Tue Oct 04 07:26:22 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.5	18976	PASS
75	95	30	60	52.3	50883	PASS
95	95	100	100	100.0	97269	PASS
96	95	5	9	6.5	6333	PASS
173	174	0.00	2	1.1	881	PASS
174	95	50	100	84.1	81803	PASS
175	174	5	9	7.4	6055	PASS
176	174	95	101	95.6	78203	PASS
177	176	5	9	6.6	5185	PASS