

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100622\
 Data File : VY010850.D
 Acq On : 06 Oct 2022 17:15
 Operator : KP/MD
 Sample : N5004-05RE
 Misc : 6.33g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D22010-18RE

Quant Time: Oct 07 02:30:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	213881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	369059	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	218874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	40010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	122740	56.616	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.240%
35) Dibromofluoromethane	7.716	113	123451	50.391	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.780%
50) Toluene-d8	10.179	98	289065	34.502	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	69.000%
62) 4-Bromofluorobenzene	12.477	95	56243	14.576	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	29.160%
Target Compounds						
16) Acetone	3.936	43	10713	19.457	ug/l	97
19) Methyl tert-butyl Ether	5.216	73	39488	7.460	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100622\
Data File : VY010850.D
Acq On : 06 Oct 2022 17:15
Operator : KP/MD
Sample : N5004-05RE
Misc : 6.33g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D22010-18RE

Quant Time: Oct 07 02:30:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

