

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100622\
 Data File : VY010854.D
 Acq On : 06 Oct 2022 18:46
 Operator : KP/MD
 Sample : N5004-10RE
 Misc : 6.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D22008-18RE

Quant Time: Oct 07 02:31:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	168869	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	283369	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	151083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	25381	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	108225	63.227	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	126.460%	
35) Dibromofluoromethane	7.710	113	109282	58.097	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	116.200%	
50) Toluene-d8	10.179	98	231468	35.982	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	71.960%	
62) 4-Bromofluorobenzene	12.477	95	37257	12.030	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	24.060%#	
Target Compounds						
16) Acetone	3.948	43	7235	16.643	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100622\
Data File : VY010854.D
Acq On : 06 Oct 2022 18:46
Operator : KP/MD
Sample : N5004-10RE
Misc : 6.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D22008-18RE

Quant Time: Oct 07 02:31:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

