

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101322\
 Data File : VY010950.D
 Acq On : 13 Oct 2022 11:06
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 14 02:09:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.346	0.355	-2.6	97	0.00
3 P	Chloromethane	0.576	0.522	9.4	88	0.00
4 C	Vinyl Chloride	0.694	0.605	12.8#	83	0.00
5 T	Bromomethane	0.492	0.398	19.1	82	0.00
6 T	Chloroethane	0.462	0.399	13.6	83	0.00
7 T	Trichlorofluoromethane	0.845	0.846	-0.1	98	0.00
8 T	Diethyl Ether	0.272	0.299	-9.9	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.521	-7.6	105	0.00
10 T	Methyl Iodide	0.648	0.661	-2.0	94	0.00
11 T	Tert butyl alcohol	0.071	0.046	35.2#	98	0.00
12 CM	1,1-Dichloroethene	0.473	0.497	-5.1#	102	0.00
13 T	Acrolein	0.039	0.025	35.9#	72	0.00
14 T	Allyl chloride	0.708	0.815	-15.1	110	0.00
15 T	Acrylonitrile	0.132	0.155	-17.4	115	0.00
16 T	Acetone	0.129	0.147	-14.0	101	0.00
17 T	Carbon Disulfide	1.528	1.466	4.1	93	0.00
18 T	Methyl Acetate	0.329	0.372	-13.1	117	0.00
19 T	Methyl tert-butyl Ether	1.237	1.363	-10.2	106	0.00
20 T	Methylene Chloride	0.678	0.604	10.9	102	0.00
21 T	trans-1,2-Dichloroethene	0.552	0.578	-4.7	102	0.00
22 T	Diisopropyl ether	1.552	1.785	-15.0	109	0.00
23 T	Vinyl Acetate	0.940	1.080	-14.9	107	0.00
24 P	1,1-Dichloroethane	0.945	1.059	-12.1	109	0.00
25 T	2-Butanone	0.185	0.204	-10.3	102	0.00
26 T	2,2-Dichloropropane	0.897	0.926	-3.2	102	0.00
27 T	cis-1,2-Dichloroethene	0.613	0.667	-8.8	104	0.00
28 T	Bromochloromethane	0.300	0.401	-33.7#	107	0.00
29 T	Tetrahydrofuran	0.109	0.126	-15.6	112	0.00
30 C	Chloroform	0.975	1.070	-9.7#	106	0.00
31 T	Cyclohexane	0.907	0.896	1.2	99	0.00
32 T	1,1,1-Trichloroethane	0.879	0.935	-6.4	102	0.00
33 S	1,2-Dichloroethane-d4	0.507	0.486	4.1	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.332	0.302	9.0	95	0.00
36 T	1,1-Dichloropropene	0.479	0.500	-4.4	100	0.00
37 T	Ethyl Acetate	0.235	0.267	-13.6	110	0.00
38 T	Carbon Tetrachloride	0.499	0.520	-4.2	101	0.00
39 T	Methylcyclohexane	0.589	0.597	-1.4	95	0.00
40 TM	Benzene	1.380	1.472	-6.7	103	0.00
41 T	Methacrylonitrile	0.127	0.131	-3.1	102	0.00
42 TM	1,2-Dichloroethane	0.390	0.409	-4.9	102	0.00
43 T	Isopropyl Acetate	0.426	0.466	-9.4	106	0.00
44 TM	Trichloroethene	0.374	0.389	-4.0	100	0.00
45 C	1,2-Dichloropropane	0.333	0.377	-13.2#	109	0.00
46 T	Dibromomethane	0.190	0.206	-8.4	106	0.00
47 T	Bromodichloromethane	0.459	0.508	-10.7	107	0.00
48 T	Methyl methacrylate	0.200	0.224	-12.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101322\
 Data File : VY010950.D
 Acq On : 13 Oct 2022 11:06
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 14 02:09:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	109	0.00
50 S	Toluene-d8	1.135	1.140	-0.4	89	0.00
51 T	4-Methyl-2-Pentanone	0.233	0.263	-12.9	107	0.00
52 CM	Toluene	0.878	0.933	-6.3#	101	0.00
53 T	t-1,3-Dichloropropene	0.476	0.514	-8.0	102	0.00
54 T	cis-1,3-Dichloropropene	0.545	0.601	-10.3	106	0.00
55 T	1,1,2-Trichloroethane	0.267	0.292	-9.4	107	0.00
56 T	Ethyl methacrylate	0.343	0.390	-13.7	106	0.00
57 T	1,3-Dichloropropane	0.454	0.502	-10.6	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.152	0.205	-34.9#	114	0.00
59 T	2-Hexanone	0.161	0.186	-15.5	103	0.00
60 T	Dibromochloromethane	0.320	0.352	-10.0	105	0.00
61 T	1,2-Dibromoethane	0.255	0.273	-7.1	105	0.00
62 S	4-Bromofluorobenzene	0.481	0.393	18.3	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.396	0.412	-4.0	100	0.00
65 PM	Chlorobenzene	1.035	1.104	-6.7	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.378	0.409	-8.2	102	0.00
67 C	Ethyl Benzene	1.863	2.018	-8.3#	101	0.00
68 T	m/p-Xylenes	0.729	0.773	-6.0	99	0.00
69 T	o-Xylene	0.682	0.732	-7.3	100	0.00
70 T	Styrene	1.149	1.249	-8.7	100	0.00
71 P	Bromoform	0.225	0.245	-8.9	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.676	3.972	-8.1	100	0.00
74 T	N-amyl acetate	0.845	0.958	-13.4	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.687	0.769	-11.9	109	0.00
76 T	1,2,3-Trichloropropane	0.506	0.574	-13.4	107	0.00
77 T	Bromobenzene	0.843	0.893	-5.9	100	0.00
78 T	n-propylbenzene	4.477	4.846	-8.2	100	0.00
79 T	2-Chlorotoluene	2.505	2.696	-7.6	101	0.00
80 T	1,3,5-Trimethylbenzene	3.109	3.323	-6.9	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.242	0.276	-14.0	107	0.00
82 T	4-Chlorotoluene	2.619	2.803	-7.0	100	0.00
83 T	tert-Butylbenzene	2.650	2.850	-7.5	99	0.00
84 T	1,2,4-Trimethylbenzene	3.057	3.291	-7.7	98	0.00
85 T	sec-Butylbenzene	3.990	4.310	-8.0	98	0.00
86 T	p-Isopropyltoluene	3.347	3.587	-7.2	97	0.00
87 T	1,3-Dichlorobenzene	1.711	1.789	-4.6	98	0.00
88 T	1,4-Dichlorobenzene	1.695	1.764	-4.1	98	0.00
89 T	n-Butylbenzene	3.080	3.359	-9.1	98	0.00
90 T	Hexachloroethane	0.640	0.688	-7.5	102	0.00
91 T	1,2-Dichlorobenzene	1.492	1.585	-6.2	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.110	0.119	-8.2	104	0.00
93 T	1,2,4-Trichlorobenzene	0.875	0.924	-5.6	100	0.00
94 T	Hexachlorobutadiene	0.540	0.548	-1.5	98	0.00
95 T	Naphthalene	1.668	1.868	-12.0	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101322\
Data File : VY010950.D
Acq On : 13 Oct 2022 11:06
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 14 02:09:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.746	0.794	-6.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6