

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
 Data File : VY011038.D
 Acq On : 21 Oct 2022 10:00
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:20:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	291968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	460716	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	407747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	197848	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	122077	45.391	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.780%
35) Dibromofluoromethane	7.722	113	128639	50.094	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.180%
50) Toluene-d8	10.179	98	444872	46.246	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.500%
62) 4-Bromofluorobenzene	12.483	95	168380	49.375	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.740%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.906	85	89205	46.061	ug/l	99
3) Chloromethane	2.119	50	122546	44.106	ug/l	95
4) Vinyl Chloride	2.253	62	154447	49.044	ug/l	99
5) Bromomethane	2.656	94	106847	49.054	ug/l	98
6) Chloroethane	2.796	64	107237	50.957	ug/l	99
7) Trichlorofluoromethane	3.131	101	239159	51.885	ug/l	92
8) Diethyl Ether	3.527	74	82282	47.836	ug/l	81
9) 1,1,2-Trichlorotrifluo...	3.905	101	149128	51.579	ug/l	95
10) Methyl Iodide	4.094	142	186622	52.750	ug/l	90
11) Tert butyl alcohol	4.948	59	71744	235.723	ug/l	97
12) 1,1-Dichloroethene	3.875	96	136203	49.177	ug/l	87
13) Acrolein	3.729	56	73827	170.276	ug/l	97
14) Allyl chloride	4.485	41	216668	49.331	ug/l #	89
15) Acrylonitrile	5.161	53	211168	235.520	ug/l	97
16) Acetone	3.948	43	197731	253.173	ug/l #	86
17) Carbon Disulfide	4.198	76	332300	40.925	ug/l	100
18) Methyl Acetate	4.478	43	98635	44.588	ug/l #	87
19) Methyl tert-butyl Ether	5.222	73	407850	52.704	ug/l	97
20) Methylene Chloride	4.716	84	164392	48.054	ug/l #	84
21) trans-1,2-Dichloroethene	5.222	96	155436	49.711	ug/l	88
22) Diisopropyl ether	6.118	45	496278	52.758	ug/l #	92
23) Vinyl Acetate	6.064	43	1476261	262.445	ug/l #	92
24) 1,1-Dichloroethane	6.021	63	296092	52.134	ug/l	98
25) 2-Butanone	6.984	43	277715	234.474	ug/l #	87
26) 2,2-Dichloropropane	6.984	77	273553	54.240	ug/l	95
27) cis-1,2-Dichloroethene	6.984	96	188768	52.383	ug/l	89
28) Bromochloromethane	7.338	49	105616	52.305	ug/l	84
29) Tetrahydrofuran	7.344	42	167262	229.532	ug/l #	85
30) Chloroform	7.508	83	307108	53.642	ug/l	99
31) Cyclohexane	7.789	56	236070	45.870	ug/l	92
32) 1,1,1-Trichloroethane	7.703	97	270528	53.757	ug/l	96
36) 1,1-Dichloropropene	7.917	75	221393	51.621	ug/l	98
37) Ethyl Acetate	7.069	43	116941	49.249	ug/l #	94
38) Carbon Tetrachloride	7.905	117	239288	54.049	ug/l	99
39) Methylcyclohexane	9.185	83	264123	51.417	ug/l	92
40) Benzene	8.161	78	652001	51.812	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
 Data File : VY011038.D
 Acq On : 21 Oct 2022 10:00
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:20:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.344	41	156632	121.726	ug/l #	1
42) 1,2-Dichloroethane	8.234	62	184452	52.900	ug/l	91
43) Isopropyl Acetate	8.277	43	214960	50.668	ug/l #	88
44) Trichloroethene	8.941	130	178928	52.469	ug/l	96
45) 1,2-Dichloropropane	9.215	63	167194	52.365	ug/l	95
46) Dibromomethane	9.307	93	93144	52.305	ug/l	97
47) Bromodichloromethane	9.496	83	234894	54.736	ug/l	99
48) Methyl methacrylate	9.295	41	96769	50.966	ug/l #	80
49) 1,4-Dioxane	9.295	88	24509	957.898	ug/l #	82
51) 4-Methyl-2-Pentanone	10.069	43	584154	246.895	ug/l #	88
52) Toluene	10.246	92	418449	53.525	ug/l	96
53) t-1,3-Dichloropropene	10.465	75	239565	53.736	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	275090	53.921	ug/l #	83
55) 1,1,2-Trichloroethane	10.648	97	136121	52.985	ug/l	98
56) Ethyl methacrylate	10.508	69	184088	53.952	ug/l #	77
57) 1,3-Dichloropropane	10.788	76	231464	52.402	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	445742	263.080	ug/l	96
59) 2-Hexanone	10.831	43	411857	253.413	ug/l	86
60) Dibromochloromethane	10.989	129	164827	53.956	ug/l	99
61) 1,2-Dibromoethane	11.093	107	127006	51.625	ug/l	100
64) Tetrachloroethene	10.721	164	170167	53.611	ug/l	97
65) Chlorobenzene	11.520	112	455103	54.365	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.593	131	172876	55.794	ug/l	97
67) Ethyl Benzene	11.593	91	824792	55.692	ug/l	99
68) m/p-Xylenes	11.703	106	631581	110.861	ug/l	94
69) o-Xylene	12.032	106	303751	56.150	ug/l	94
70) Styrene	12.044	104	516724	56.372	ug/l	96
71) Bromoform	12.209	173	104510	54.653	ug/l #	99
73) Isopropylbenzene	12.331	105	824112	57.333	ug/l	100
74) N-amyl acetate	12.142	43	192483	52.779	ug/l #	85
75) 1,1,2,2-Tetrachloroethane	12.581	83	156064	51.734	ug/l	98
76) 1,2,3-Trichloropropane	12.636	75	205409	91.447	ug/l #	100
77) Bromobenzene	12.611	156	184618	54.947	ug/l	98
78) n-propylbenzene	12.672	91	992307	56.740	ug/l	99
79) 2-Chlorotoluene	12.757	91	550056	55.849	ug/l	98
80) 1,3,5-Trimethylbenzene	12.812	105	685520	57.159	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.379	75	55936	53.558	ug/l #	84
82) 4-Chlorotoluene	12.855	91	562386	54.842	ug/l	98
83) tert-Butylbenzene	13.074	119	618066	59.596	ug/l	93
84) 1,2,4-Trimethylbenzene	13.123	105	675622	57.129	ug/l	98
85) sec-Butylbenzene	13.257	105	897073	57.563	ug/l	99
86) p-Isopropyltoluene	13.373	119	748063	57.862	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	373567	55.710	ug/l	100
88) 1,4-Dichlorobenzene	13.446	146	365703	54.850	ug/l	99
89) n-Butylbenzene	13.696	91	707291	58.546	ug/l	99
90) Hexachloroethane	13.965	117	143485	56.237	ug/l	90
91) 1,2-Dichlorobenzene	13.739	146	330554	55.097	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	24419	48.958	ug/l	80
93) 1,2,4-Trichlorobenzene	15.007	180	202043	56.348	ug/l	99
94) Hexachlorobutadiene	15.111	225	119906	56.813	ug/l	99
95) Naphthalene	15.239	128	405149	55.483	ug/l	98
96) 1,2,3-Trichlorobenzene	15.428	180	174814	55.885	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
Data File : VY011038.D
Acq On : 21 Oct 2022 10:00
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Quant Time: Oct 22 01:20:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
 Data File : VY011038.D
 Acq On : 21 Oct 2022 10:00
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:20:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 291968
Ion Ratio Lower Upper
168 100
99 54.6 46.9 70.3




#2
Dichlorodifluoromethane
Concen: 46.061 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 85 Resp: 89205
Ion Ratio Lower Upper
85 100
87 31.3 16.1 48.2



Abundance Scan 48 (2.113 min): VY010969.D\data.ms (-40)



#3

Chloromethane

Concen: 44.106 ug/l

RT: 2.119 min Scan# 49

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050


Tgt Ion: 50 Resp: 122546

Ion Ratio Lower Upper


50 100

52 33.5 24.6 37.0


Abundance Scan 49 (2.119 min): VY011038.D\data.ms




Abundance



Abundance Scan 49 (2.119 min): VY011038.D\data.ms (-1)



Abundance Scan 70 (2.247 min): VY010969.D\data.ms (-61)



#4

Vinyl Chloride

Concen: 49.044 ug/l

RT: 2.253 min Scan# 71

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00


Tgt Ion: 62 Resp: 154447

Ion Ratio Lower Upper


62 100

64 32.0 25.0 37.6


Abundance Scan 71 (2.253 min): VY011038.D\data.ms




Abundance



Abundance Scan 71 (2.253 min): VY011038.D\data.ms (-21)









#5
Bromomethane
Concen: 49.054 ug/l
RT: 2.656 min Scan# 11
Delta R.T. 0.012 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 106847
Ion Ratio Lower Upper
94 100
96 89.8 73.5 110.3



#6
Chloroethane
Concen: 50.957 ug/l
RT: 2.796 min Scan# 160
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 64 Resp: 107237
Ion Ratio Lower Upper
64 100
66 32.1 25.4 38.0









#7
Trichlorofluoromethane
Concen: 51.885 ug/l
RT: 3.131 min Scan# 213
Delta R.T. 0.012 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 239159
Ion Ratio Lower Upper
101 100
103 65.9 48.0 72.0



#8
Diethyl Ether
Concen: 47.836 ug/l
RT: 3.527 min Scan# 280
Delta R.T. -0.001 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 74 Resp: 82282
Ion Ratio Lower Upper
74 100
45 89.9 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.579 ug/l
RT: 3.905 min Scan# 34
Delta R.T. 0.012 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 149128
Ion Ratio Lower Upper
101 100
85 44.5 40.6 60.8
151 78.5 64.9 97.3



#10
Methyl Iodide
Concen: 52.750 ug/l
RT: 4.094 min Scan# 373
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:142 Resp: 186622
Ion Ratio Lower Upper
142 100
127 43.5 41.5 62.3
141 14.3 11.9 17.9





#11

Tert butyl alcohol

Concen: 235.723 ug/l

RT: 4.948 min Scan# 514

Delta R.T. -0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion: 59 Resp: 71744

Ion Ratio Lower Upper

59 100

57 9.2 6.4 9.6



#12

1,1-Dichloroethene

Concen: 49.177 ug/l

RT: 3.875 min Scan# 337

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00




Tgt Ion: 96 Resp: 136203


Ion Ratio Lower Upper

96 100

61 157.5 145.8 218.6

98 64.8 52.5 78.7









#13
Acrolein
Concen: 170.276 ug/l
RT: 3.729 min Scan# 312
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 73827
Ion Ratio Lower Upper
56 100
55 69.9 57.6 86.4



#14
Allyl chloride
Concen: 49.331 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.012 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 41 Resp: 216668
Ion Ratio Lower Upper
41 100
39 67.1 58.8 88.2
76 41.9 25.1 37.7#









#15
Acrylonitrile
Concen: 235.520 ug/l
RT: 5.161 min Scan# 54
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 211168
Ion Ratio Lower Upper
53 100
52 81.1 67.0 100.4
51 36.0 29.8 44.6



#16
Acetone
Concen: 253.173 ug/l
RT: 3.948 min Scan# 349
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 43 Resp: 197731
Ion Ratio Lower Upper
43 100
58 34.9 22.0 33.0#









#17
Carbon Disulfide
Concen: 40.925 ug/l
RT: 4.198 min Scan# 390
Delta R.T. 0.012 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 332300
Ion Ratio Lower Upper
76 100
78 9.1 7.4 11.0



#18
Methyl Acetate
Concen: 44.588 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.005 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 43 Resp: 98635
Ion Ratio Lower Upper
43 100
74 26.6 16.6 24.8#









#19
Methyl tert-butyl Ether
Concen: 52.704 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 407850
Ion Ratio Lower Upper
73 100
57 21.9 18.6 27.8



#20
Methylene Chloride
Concen: 48.054 ug/l
RT: 4.716 min Scan# 475
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 84 Resp: 164392
Ion Ratio Lower Upper
84 100
49 114.0 114.4 171.6#
51 37.1 35.3 52.9
86 64.2 53.7 80.5









#21
trans-1,2-Dichloroethene
Concen: 49.711 ug/l
RT: 5.222 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 155436
Ion Ratio Lower Upper
96 100
61 131.8 120.5 180.7
98 65.2 49.8 74.8



#22
Diisopropyl ether
Concen: 52.758 ug/l
RT: 6.118 min Scan# 705
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 45 Resp: 496278
Ion Ratio Lower Upper
45 100
43 56.0 49.3 73.9
87 29.5 19.2 28.8#
59 12.2 8.4 12.6









#23
 Vinyl Acetate
 Concen: 262.445 ug/l
 RT: 6.064 min Scan# 695
 Delta R.T. 0.006 min
 Lab File: VY011038.D
 Acq: 21 Oct 2022 10:00




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 1476261
 Ion Ratio Lower Upper
 43 100
 86 11.7 7.1 10.7#



#24
 1,1-Dichloroethane
 Concen: 52.134 ug/l
 RT: 6.021 min Scan# 689
 Delta R.T. 0.006 min
 Lab File: VY011038.D
 Acq: 21 Oct 2022 10:00

Tgt Ion: 63 Resp: 296092
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.6 10.8
 100 4.7 2.0 5.8









#25
2-Butanone
Concen: 234.474 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 277715
Ion Ratio Lower Upper
43 100
72 28.3 17.7 26.5#



#26
2,2-Dichloropropane
Concen: 54.240 ug/l
RT: 6.984 min Scan# 847
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 77 Resp: 273553
Ion Ratio Lower Upper
77 100
97 23.0 10.4 31.4









#27
cis-1,2-Dichloroethene
Concen: 52.383 ug/l
RT: 6.984 min Scan# 846
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 188768
Ion Ratio Lower Upper
96 100
61 141.0 0.0 319.8
98 65.0 0.0 127.0



#28
Bromochloromethane
Concen: 52.305 ug/l
RT: 7.338 min Scan# 905
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 49 Resp: 105616
Ion Ratio Lower Upper
49 100
129 2.1 0.0 2.2
130 82.1 55.4 83.0





#29

Tetrahydrofuran

Concen: 229.532 ug/l

RT: 7.344 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 167262

Ion Ratio Lower Upper

42 100

72 47.7 30.3 45.5#

71 44.3 29.0 43.4#



#30

Chloroform

Concen: 53.642 ug/l

RT: 7.508 min Scan# 933

Delta R.T. 0.006 min

Lab File: VY011038.D




Acq: 21 Oct 2022 10:00


Tgt Ion: 83 Resp: 307108

Ion Ratio Lower Upper

83 100

85 65.7 51.8 77.6









#31
Cyclohexane
Concen: 45.870 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 236070
Ion Ratio Lower Upper
56 100
69 35.1 26.7 40.1
84 93.4 67.2 100.8



#32
1,1,1-Trichloroethane
Concen: 53.757 ug/l
RT: 7.703 min Scan# 965
Delta R.T. 0.005 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 97 Resp: 270528
Ion Ratio Lower Upper
97 100
99 64.3 50.7 76.1
61 41.3 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 45.391 ug/l

RT: 8.142 min Scan# 110

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 122077

Ion Ratio Lower Upper

65 100

67 53.6 0.0 99.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00




Tgt Ion: 114 Resp: 460716


Ion Ratio Lower Upper

114 100

63 19.0 0.0 43.2

88 15.3 0.0 31.8









#35
Dibromofluoromethane
Concen: 50.094 ug/l
RT: 7.722 min Scan# 90
Delta R.T. 0.012 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 128639
Ion Ratio Lower Upper
113 100
111 101.7 82.6 124.0
192 19.7 17.6 26.4



#36
1,1-Dichloropropene
Concen: 51.621 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 75 Resp: 221393
Ion Ratio Lower Upper
75 100
110 36.0 17.1 51.2
77 30.9 24.4 36.6





#37

Ethyl Acetate

Concen: 49.249 ug/l

RT: 7.069 min Scan# 8

Delta R.T. -0.001 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 116941

Ion Ratio Lower Upper

43 100

61 14.5 9.7 14.5#

70 11.3 7.1 10.7#



#38

Carbon Tetrachloride

Concen: 54.049 ug/l

RT: 7.905 min Scan# 998

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00




Tgt Ion: 117 Resp: 239288


Ion Ratio Lower Upper

117 100

119 97.0 77.9 116.9

121 30.4 24.8 37.2





#39

Methylcyclohexane

Concen: 51.417 ug/l

RT: 9.185 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 264123

Ion Ratio Lower Upper

83 100

55 73.4 66.5 99.7

98 49.3 37.9 56.9



#40

Benzene

Concen: 51.812 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. 0.006 min

Lab File: VY011038.D




Acq: 21 Oct 2022 10:00


Tgt Ion: 78 Resp: 652001

Ion Ratio Lower Upper

78 100

77 23.4 20.4 30.6









#41
Methacrylonitrile
Concen: 121.726 ug/l
RT: 7.344 min Scan# 906
Delta R.T. 0.037 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 41 Resp: 156632
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3#
67 0.0 0.0 0.0
52 0.0 0.0 0.0



#42
1,2-Dichloroethane
Concen: 52.900 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 62 Resp: 184452
Ion Ratio Lower Upper
62 100
98 10.4 0.0 14.8





#43

Isopropyl Acetate

Concen: 50.668 ug/l

RT: 8.277 min Scan# 1105

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 214960

Ion Ratio Lower Upper

43 100

61 31.4 20.2 30.2#

87 15.0 8.9 13.3#



#44

Trichloroethene

Concen: 52.469 ug/l

RT: 8.941 min Scan# 1168

Delta R.T. 0.006 min

Lab File: VY011038.D




Acq: 21 Oct 2022 10:00


Tgt Ion:130 Resp: 178928

Ion Ratio Lower Upper

130 100

95 98.3 0.0 188.2





#45

1,2-Dichloropropane

Concen: 52.365 ug/l

RT: 9.215 min Scan# 11

Delta R.T. -0.001 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 167194

Ion Ratio Lower Upper

63 100

65 31.5 23.0 34.6



#46

Dibromomethane

Concen: 52.305 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.006 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00




Tgt Ion: 93 Resp: 93144


Ion Ratio Lower Upper

93 100

95 82.4 64.3 96.5

174 100.5 83.4 125.0









#47
Bromodichloromethane
Concen: 54.736 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 234894
Ion Ratio Lower Upper
83 100
85 63.9 51.9 77.9
127 8.4 6.6 10.0



#48
Methyl methacrylate
Concen: 50.966 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 41 Resp: 96769
Ion Ratio Lower Upper
41 100
69 97.9 58.2 87.2#
39 54.1 47.4 71.0





#49

1,4-Dioxane

Concen: 957.898 ug/l

RT: 9.295 min Scan# 1226

Delta R.T. -0.000 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 24509

Ion Ratio Lower Upper

88 100

43 28.0 32.8 49.2#

58 60.7 59.8 89.6



#50

Toluene-d8

Concen: 46.246 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. -0.000 min

Lab File: VY011038.D




Acq: 21 Oct 2022 10:00


Tgt Ion: 98 Resp: 444872

Ion Ratio Lower Upper

98 100

100 65.1 50.2 75.2









#51
4-Methyl-2-Pentanone
Concen: 246.895 ug/l
RT: 10.069 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 584154
Ion Ratio Lower Upper
43 100
58 42.1 28.1 42.1#



#52
Toluene
Concen: 53.525 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 92 Resp: 418449
Ion Ratio Lower Upper
92 100
91 172.4 142.6 213.8









#53
t-1,3-Dichloropropene
Concen: 53.736 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 239565
Ion Ratio Lower Upper
75 100
77 31.4 25.4 38.2



#54
cis-1,3-Dichloropropene
Concen: 53.921 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 75 Resp: 275090
Ion Ratio Lower Upper
75 100
77 31.6 25.6 38.4
39 40.5 47.6 71.4#









#55
1,1,2-Trichloroethane
Concen: 52.985 ug/l
RT: 10.648 min Scan# 1447
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 136121




Ion	Ratio	Lower	Upper
97	100		
83	84.9	69.8	104.8
85	55.4	44.3	66.5
99	63.4	49.8	74.6




#56
Ethyl methacrylate
Concen: 53.952 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 69 Resp: 184088

Ion	Ratio	Lower	Upper
69	100		
41	59.2	63.4	95.0#
39	30.1	35.8	53.8#





#57

1,3-Dichloropropane

Concen: 52.402 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 231464

Ion Ratio Lower Upper

76 100

78 32.3 25.0 37.6



#58

2-Chloroethyl Vinyl ether

Concen: 263.080 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. -0.001 min

Lab File: VY011038.D




Acq: 21 Oct 2022 10:00


Tgt Ion: 63 Resp: 445742

Ion Ratio Lower Upper

63 100

106 28.9 21.5 32.3









#59
2-Hexanone
Concen: 253.413 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 411857
Ion Ratio Lower Upper
43 100
58 58.6 24.4 73.4



#60
Dibromochloromethane
Concen: 53.956 ug/l
RT: 10.989 min Scan# 1504
Delta R.T. 0.005 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 129 Resp: 164827
Ion Ratio Lower Upper
129 100
127 78.3 38.7 116.1









#61
1,2-Dibromoethane
Concen: 51.625 ug/l
RT: 11.093 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 127006
Ion Ratio Lower Upper
107 100
109 93.9 75.4 113.2



#62
4-Bromofluorobenzene
Concen: 49.375 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 95 Resp: 168380
Ion Ratio Lower Upper
95 100
174 82.9 0.0 177.0
176 80.3 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.495 min Scan# 11
Delta R.T. 0.005 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 407747
Ion Ratio Lower Upper
117 100
82 52.3 43.0 64.4
119 31.8 25.4 38.0



#64
Tetrachloroethene
Concen: 53.611 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:164 Resp: 170167
Ion Ratio Lower Upper
164 100
166 128.9 100.4 150.6
129 95.0 71.4 107.0
131 89.7 70.8 106.2









#65
Chlorobenzene
Concen: 54.365 ug/l
RT: 11.520 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 455103
Ion Ratio Lower Upper
112 100
114 32.3 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 55.794 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:131 Resp: 172876
Ion Ratio Lower Upper
131 100
133 95.3 47.9 143.6
119 66.4 30.8 92.3









#67
Ethyl Benzene
Concen: 55.692 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 824792
Ion Ratio Lower Upper
91 100
106 31.4 24.5 36.7



#68
m/p-Xylenes
Concen: 110.861 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 106 Resp: 631581
Ion Ratio Lower Upper
106 100
91 198.8 166.6 249.8









#69
o-Xylene
Concen: 56.150 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 303751
Ion Ratio Lower Upper
106 100
91 211.1 110.3 331.0



#70
Styrene
Concen: 56.372 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:104 Resp: 516724
Ion Ratio Lower Upper
104 100
78 47.7 42.4 63.6
103 54.5 44.4 66.6









#71
Bromoform
Concen: 54.653 ug/l
RT: 12.209 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 104510
Ion Ratio Lower Upper
173 100
175 48.8 24.6 74.0
254 0.2 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:152 Resp: 197848
Ion Ratio Lower Upper
152 100
115 58.2 28.2 84.7
150 177.7 0.0 347.6









#73
Isopropylbenzene
Concen: 57.333 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 824112
Ion Ratio Lower Upper
105 100
120 26.8 13.4 40.2



#74
N-ethyl acetate
Concen: 52.779 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 43 Resp: 192483
Ion Ratio Lower Upper
43 100
70 47.6 28.7 43.1#
55 26.2 17.3 25.9#
61 26.3 17.0 25.6#









#75
1,1,2,2-Tetrachloroethane
Concen: 51.734 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 156064
Ion Ratio Lower Upper
83 100
131 10.1 5.6 16.8
85 63.1 32.3 96.8



#76
1,2,3-Trichloropropane
Concen: 91.447 ug/l
RT: 12.636 min Scan# 1774
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 75 Resp: 205409
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 54.947 ug/l
RT: 12.611 min Scan# 1769
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 184618
Ion Ratio Lower Upper
156 100
77 186.9 95.5 286.4
158 97.8 48.0 144.0



#78
n-propylbenzene
Concen: 56.740 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 91 Resp: 992307
Ion Ratio Lower Upper
91 100
120 23.1 11.3 33.9









#79
2-Chlorotoluene
Concen: 55.849 ug/l
RT: 12.757 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 550056
Ion Ratio Lower Upper
91 100
126 34.7 16.9 50.6



#80
1,3,5-Trimethylbenzene
Concen: 57.159 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 105 Resp: 685520
Ion Ratio Lower Upper
105 100
120 49.3 24.4 73.4









#81
trans-1,4-Dichloro-2-butene
Concen: 53.558 ug/l
RT: 12.379 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 55936
Ion Ratio Lower Upper
75 100
53 84.6 85.3 127.9#
89 43.8 37.0 55.4



#82
4-Chlorotoluene
Concen: 54.842 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 91 Resp: 562386
Ion Ratio Lower Upper
91 100
126 34.9 17.0 51.0









#83
tert-Butylbenzene
Concen: 59.596 ug/l
RT: 13.074 min Scan# 119
Delta R.T. -0.001 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 618066
Ion Ratio Lower Upper
119 100
91 61.8 34.8 104.4
134 25.6 13.3 39.9



#84
1,2,4-Trimethylbenzene
Concen: 57.129 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:105 Resp: 675622
Ion Ratio Lower Upper
105 100
120 45.9 22.3 66.9









#85
 sec-Butylbenzene
 Concen: 57.563 ug/l
 RT: 13.257 min Scan# 1875
 Delta R.T. 0.006 min
 Lab File: VY011038.D
 Acq: 21 Oct 2022 10:00




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 897073
 Ion Ratio Lower Upper
 105 100
 134 20.7 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 57.862 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.006 min
 Lab File: VY011038.D
 Acq: 21 Oct 2022 10:00

Tgt Ion:119 Resp: 748063
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.2 39.5
 91 22.5 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 55.710 ug/l

RT: 13.367 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 373567

Ion Ratio Lower Upper

146 100

111 39.4 19.5 58.5

148 63.0 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 54.850 ug/l

RT: 13.446 min Scan# 1907

Delta R.T. -0.001 min

Lab File: VY011038.D

Acq: 21 Oct 2022 10:00




Tgt Ion:146 Resp: 365703


Ion Ratio Lower Upper

146 100

111 38.6 19.3 57.8

148 63.9 32.6 97.8









#89
n-Butylbenzene
Concen: 58.546 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 707291
Ion Ratio Lower Upper
91 100
92 53.7 26.3 78.9
134 26.2 13.0 38.9



#90
Hexachloroethane
Concen: 56.237 ug/l
RT: 13.965 min Scan# 1992
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 117 Resp: 143485
Ion Ratio Lower Upper
117 100
201 76.6 43.0 128.8









#91
1,2-Dichlorobenzene
Concen: 55.097 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 330554
Ion Ratio Lower Upper
146 100
111 41.2 20.3 60.9
148 63.5 32.3 96.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 48.958 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion: 75 Resp: 24419
Ion Ratio Lower Upper
75 100
155 91.0 38.0 113.9
157 116.0 47.8 143.3









#93
1,2,4-Trichlorobenzene
Concen: 56.348 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 202043
Ion Ratio Lower Upper
180 100
182 95.7 47.3 141.9
145 30.5 14.6 43.8



#94
Hexachlorobutadiene
Concen: 56.813 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. -0.000 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:225 Resp: 119906
Ion Ratio Lower Upper
225 100
223 63.3 30.8 92.3
227 63.6 31.6 94.7









#95
Naphthalene
Concen: 55.483 ug/l
RT: 15.239 min Scan# 21
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 405149
Ion Ratio Lower Upper
128 100
127 12.5 11.0 16.4
129 11.1 8.4 12.6



#96
1,2,3-Trichlorobenzene
Concen: 55.885 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. 0.006 min
Lab File: VY011038.D
Acq: 21 Oct 2022 10:00

Tgt Ion:180 Resp: 174814
Ion Ratio Lower Upper
180 100
182 94.7 47.3 141.9
145 32.2 15.3 45.9

