

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
 Data File : VY011048.D
 Acq On : 21 Oct 2022 14:22
 Operator : KP/MD
 Sample : N5217-11
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-M1-3.5-4.0

Quant Time: Oct 22 01:23:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	327967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	549390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	482094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	221772	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	176890	58.552	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.100%
35) Dibromofluoromethane	7.716	113	186490	61.386	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.780%
50) Toluene-d8	10.179	98	674295	58.781	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	117.560%
62) 4-Bromofluorobenzene	12.477	95	226945	56.233	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
Data File : VY011048.D
Acq On : 21 Oct 2022 14:22
Operator : KP/MD
Sample : N5217-11
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-M1-3.5-4.0

Quant Time: Oct 22 01:23:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

