

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
 Data File : VY011052.D
 Acq On : 21 Oct 2022 15:53
 Operator : KP/MD
 Sample : N5217-19
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-E1-7.5-8.0

Quant Time: Oct 22 01:23:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	261978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	448311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	396720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	162041	67.148	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 134.300%			
35) Dibromofluoromethane	7.716	113	169394	68.586	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 137.180%			
50) Toluene-d8	10.179	98	598827	63.972	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 127.940%			
62) 4-Bromofluorobenzene	12.477	95	205088	62.627	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 125.260%			
Target Compounds						
20) Methylene Chloride	4.710	84	38143	7.065	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
Data File : VY011052.D
Acq On : 21 Oct 2022 15:53
Operator : KP/MD
Sample : N5217-19
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-E1-7.5-8.0

Quant Time: Oct 22 01:23:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

