

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
 Data File : VY011055.D
 Acq On : 21 Oct 2022 17:01
 Operator : KP/MD
 Sample : N5217-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-A1-3.5-4.0RE

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/24/2022
 Supervised By :Mahesh Dadoda 10/24/2022

Quant Time: Oct 22 01:24:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2384	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	3750	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	2817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1647	75.000	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.000%
35) Dibromofluoromethane	7.710	113	690	32.243	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	64.480%
50) Toluene-d8	10.179	98	4501	57.484	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.960%
62) 4-Bromofluorobenzene	12.483	95	733	24.884	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	49.760%
Target Compounds						Qvalue
3) Chloromethane	2.119	50	118	5.201	ug/l	# 77
20) Methylene Chloride	4.722	84	591m	17.110	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102122\
Data File : VY011055.D
Acq On : 21 Oct 2022 17:01
Operator : KP/MD
Sample : N5217-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-A1-3.5-4.0RE

Quant Time: Oct 22 01:24:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/24/2022
Supervised By :Mahesh Dadoda 10/24/2022

