


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv102219\
Data File : VY000356.D
Acq On : 22 Oct 2019 10:03
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y101119S.M
Title : SW846 8260
Last Update : Thu Oct 17 02:03:07 2019



AutoFind: Scans 1762, 1763, 1764; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	38304	PASS
75	95	30	60	52.3	101831	PASS
95	95	100	100	100.0	194858	PASS
96	95	5	9	6.5	12742	PASS
173	174	0.00	2	0.8	1310	PASS
174	95	50	100	84.6	164928	PASS
175	174	5	9	6.9	11361	PASS
176	174	95	101	96.2	158741	PASS
177	176	5	9	6.7	10578	PASS