

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
 Data File : VY006459.D
 Acq On : 25 Oct 2021 12:18
 Operator : SY/MD
 Sample : M4315-01
 Misc : 4.53G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-D-S

Quant Time: Oct 26 02:42:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	103454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	174929	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	160167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	68431	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	72104	56.638	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.280%
35) Dibromofluoromethane	7.722	113	52170	53.507	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.020%
50) Toluene-d8	10.185	98	222459	53.823	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.640%
62) 4-Bromofluorobenzene	12.483	95	65925	47.350	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
Data File : VY006459.D
Acq On : 25 Oct 2021 12:18
Operator : SY/MD
Sample : M4315-01
Misc : 4.53G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-D-S

Quant Time: Oct 26 02:42:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

