

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
 Data File : VY006462.D
 Acq On : 25 Oct 2021 13:29
 Operator : SY/MD
 Sample : M4313-05
 Misc : 4.99G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 4743

Quant Time: Oct 26 02:43:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	99282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	169771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	157578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	66833	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	71223	58.297	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.600%
35) Dibromofluoromethane	7.722	113	50603	53.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.960%
50) Toluene-d8	10.185	98	217114	54.125	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.260%
62) 4-Bromofluorobenzene	12.483	95	64539	47.762	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
Data File : VY006462.D
Acq On : 25 Oct 2021 13:29
Operator : SY/MD
Sample : M4313-05
Misc : 4.99G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
4743

Quant Time: Oct 26 02:43:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

