

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
 Data File : VY006463.D
 Acq On : 25 Oct 2021 13:53
 Operator : SY/MD
 Sample : M4306-01
 Misc : 4.71G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-PIT-SOIL

Quant Time: Oct 26 02:43:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	100352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	171222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	160827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	70324	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	76467	61.922	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.840%
35) Dibromofluoromethane	7.722	113	52688	55.209	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.420%
50) Toluene-d8	10.185	98	219102	54.158	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.320%
62) 4-Bromofluorobenzene	12.483	95	68070	49.949	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102521\
Data File : VY006463.D
Acq On : 25 Oct 2021 13:53
Operator : SY/MD
Sample : M4306-01
Misc : 4.71G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-PIT-SOIL

Quant Time: Oct 26 02:43:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

